

鼠兔双标三色荧光检测试剂盒（温和型）

CatalogNo: RS0069

| Key Features

Applications

- IF, mIHC

| Recommended Dilution Ratios

Ready to use

| Storage

Storage*

See datasheet


| Basic Information

| Immunogen Information

| Target Information

Protein Name

| Validation Data



Fluorescence multiplex immunohistochemical analysis of Human tonsil tissue (formalin-fixed paraffin-embedded section). The immunostaining was performed on a Leica Biosystems BOND® MAX instrument with an multiple-Fluorescence kit (RS0069, Immunoway). The section was incubated in 6 rounds of staining; sequentially for Anti-antibodies; each using a separate fluorescent tyramide signal amplification system. mIHC Antibody Sprng Buffer(YS0124)was used in between rounds of tyramide signal amplification to remove the antibody from the previous round, to avoid any cross-reactivity. DAPI (dark blue) was used as a nuclear counter stain. Microscopy and pseudocoloring of individual dyes was performed using a Slideviewer Imaging System (3D histech).

| Contact information

Orders: order@immunoway.com
Support: tech@immunoway.com
Telephone: 408-747-0189 (USA) 400-8787-807(China)
Website: http://www.immunoway.com
Address: 2200 Ringwood Ave San Jose, CA 95131 USA



Please scan the QR code
to access additional
product information:
鼠兔双标三色荧光检测
试剂盒 (温和型)

For Research Use Only. Not for Use in Diagnostic Procedures.

[Antibody](#) | [ELISA Kits](#) | [Protein](#) | [Reagents](#)